

TENDER NOTICE

No:CEPC/ASIDE/1/2011

Dated: 6th February 2013

The Chief Administrator / Head on behalf of CEPC Laboratory and Technical Division, The Cashew Export Promotion Council of India (CEPC), Kollam, invites sealed tenders for the purchase of “**Autoclaves (2 nos.) & Digital Temperature monitoring equipment (1 no.)**” as per the list below. These equipments are to be installed at CEPC Laboratory and Technical Division, Cashew Bhavan, Mundakkal, Kollam, Kerala.

GENERAL CONDITIONS

The CEPC would like to have at its Laboratory, the latest technology and the highest quality testing instruments. Quotation should give detailed specifications of equipments on offer, brand name, terms & conditions of supply, guarantee/warranty period, after sales service details, delivery period and the customer list including the name of persons with telephone number/Email ID etc. An EMD as specified under should be enclosed with the tender. A statement of compliance with specifications should accompany the quotation. Onsite training of personnel to be required immediately after installation of the equipment.

Technical Specification should include Data Sheet, configurations, brand, model, essential accessories, optional accessories, spares & consumables for five years period, year of introduction of the model offered etc. along with the technical offer and should specify commercial terms and conditions of the offer.

The services offered like installation, commissioning, training etc. are to be detailed. The terms & conditions for labour only, AMC & comprehensive AMC after the warranty period etc. to be offered.

Late quotations will not be accepted.

| | |
|---|---|
| Last date & time for the receipt of quotations | 18/02/2013 (04:00 PM) |
| Date & time of opening the quotations | 19/02/2013 (10:00 AM) in the presence of the tenderer (if present)/ authorised personnel. |
| EMD : Rs. 5000/- (autoclaves- 2 nos.) Rs. 2500/- (Digital temperature -1no.) monitoring equipment) (RTD Sensor / probe) | EMD should be either by a demand draft or bank guarantee in favour of CEPC Laboratory, Kollam, Kerala. |
| Mode of Payment | After satisfactory installation & commissioning of the equipment at the site. |

| Sl. No. | Item | Specifications | Quantity |
|---------|--|---|----------|
| 1. | Autoclaves (Vertical) | <ul style="list-style-type: none"> ➤ Chamber Material- Inner, SS 304 and outer constructed heavy stainless steel/ MS duly heat cured epoxy coated equipped with durable moulded silicon gasket, castors for ease of movement. ➤ Entire unit should be supplied with Tripod stand for keeping work baskets, fitted with high quality heating element, Pressure gauge, steam cock, safety valve fitted on lid. ➤ Inner Chamber Dia. & Depth- 350 x 600mm (approx.) Capacity : 50-55 Lit. ➤ Sterilization Temperature range:- 105°-126°C or above ➤ Pressure gauge range:0- 2kg/cm² ➤ Digital Temperature Controller Cum Indicator (Microprocessor based) with a resolution 0.1°C & accuracy +/- 1°C. ➤ Power supply 220 - 240 volts AC ➤ Accessories : 1. work baskets 2. Servo based stabilizer (5KVA) ➤ Calibration certificate from an ISO 17025 ;2005 accredited laboratory ➤ Conformance to national & international standards with CE certificate. | 2 nos. |
| 2. | Digital Temperature Monitoring Equipment (RTD Thermometer/probe) | <ul style="list-style-type: none"> ➤ Temperature range: -20°C TO 550°C (approx.) ➤ Resolution:0.1°C ➤ Accuracy : +/- 0.5°C ➤ Response Time : 500 msec. ➤ Display : Backlit LCD ➤ Power Supply ; Re chargeable battery pack/Adaptor/carrying case ➤ Data logger -down load later to PC ➤ Battery Operation : Minimum 48 hrs continuous operation ➤ Accessories- General probe, Sensor on extension cable (Range 12 to 60°C), Accuracy +/- 0.1°C. | 1 no. |

TERMS AND CONDITIONS

1. **Submission of tender:-** The tender shall be duly signed, stamped & sealed, superscribed as “Tender for supply of **Autoclaves & Digital Temperature monitoring equipment** with due date 18-02-2013” and submitted to the CEPC Office at Kollam.
2. Warranty / AMC for five years to be included in the offer.
3. Onsite training to CEPC scientists to be provided after installation for smooth operation of the instrument.
4. Technical support at short notice in case of breakdown of equipment should be ensured. Replacement of imported components must not take more than three weeks.
5. Time of delivery - 30days from the date of supply order.
6. Users list - To be furnished with contact details.
7. All manuals for operation, service and maintenance with electronic circuit diagrams to be provided.
8. The offer should also stipulate pre-installation requirements and utilities if any for the instrument. The suppliers should specify clearly; their scope of supply and items that do not fall under their scope of supply.
9. Tenders received late for any reason whatsoever shall not be entertained. CEPC has the right to reject any or all tenders without assigning any reason.
10. For instruments to be imported prices may be quoted in Indian Rupee also. Our organization is not exempted from the payment of customs duty.

Sd-

Chief Administrator/Head

TENDER FORM

Tender No:.....

Last date of Receipt: 18/02/2013 (04:00PM)

Date:.....

Due date of Opening: 19/02/2013 (10:00AM)

Full name and Address of the Tenderer with Tel./Fax/E-mail ID.

From,,.....

.....

.....

To,

The Chief Administrator / Head,

CEPC Laboratory & Technical Division,

(The Cashew Export Promotion Council of India)

(Sponsored by Ministry of Commerce, Govt. of India)

CASHEW BHAVAN, Mundakkal, Kollam – 691001, Kerala, India.

Dear Sir,

I/We hereby have gone through the tendering conditions pertaining “The Purchase of Autoclaves & Digital Temperature Monitoring Equipment”; and agree to abide by your general conditions and terms and conditions of contract as per your notification dated/02/2013.

I/We hereby offer to supply the stores conforming to the tender specifications and agree to keep the price valid for your acceptance for a period of 90 days from the date of opening of the tender.

I/We am/are fully aware of the nature of the instruments required and my/our offer is to supply the instruments strictly in accordance with the requirements.

I/we certify that we are the manufacturers/ authorized dealers to supply the item. The tender has a total of..... Number of pages which includes the details of equipments and the price quoted.

Yours faithfully,

Stamp and Signature of the Tenderer